



CERTIFICATE OF ANALYSIS

Product:
GoBio™ Prep 16x600 40/10 000 SEC

Article number:
55 435 026

Lot number: D04022601

Storage solution:

20% ethanol

	Measured Value	Specification
K _D (Ovalbumin)*:	0.71	0.65-0.85
Efficiency:	7062 N/m	≥5000 N/m
Asymmetry:	1.4	0.7-1.4
Bed height:	600 mm	600-620 mm

*Property measured on resin.

Resin lot number: 157736

Release date (year-month-day): 2026-02-24

Expiry date (year-month): 2029-08

Manufacturing date (year-month): 2026-02

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2026-02-24

By Markus Norén

DocuSigned by:

Markus Norén

6853712A40184AF...

Quality Assurance