
CERTIFICATE OF ANALYSIS

Product:
GoBio™ Prep 16x600 40/1000 SEC

Article number:
55 430 026

Lot number: D05062302

Storage solution:

20% ethanol

	Measured Value	Specification
K _D (Ovalbumin)*:	0.64	0.62-0.75
Efficiency:	8280 N/m	≥5000 N/m
Asymmetry:	1.0	0.7-1.4
Bed height:	600 mm	600-620 mm

*Property measured on resin.

Resin lot number: 155025

Release date (year-month-day): 2023-06-26

Expiry date (year-month): 2026-09

Manufacturing date (year-month): 2023-06

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2023-06-26

By Philip Eriksson

DocuSigned by:

158C686C7E5E4A0...

Quality Assurance