



CERTIFICATE OF ANALYSIS

Product:
GoBio™ Prep 16x600 40/100 SEC

Article number:
55 434 026

Lot number: D06032504

Storage solution:

20% ethanol

	Measured Value	Specification
K _D (Ovalbumin)*:	0.55	0.40-0.60
Efficiency:	9525 N/m	≥5000 N/m
Asymmetry:	1.3	0.7-1.4
Bed height:	600 mm	600-620 mm

*Property measured on resin.

Resin lot number: 157833

Release date (year-month-day): 2025-03-24


Expiry date (year-month): 2029-11

Manufacturing date (year-month): 2025-03

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2025-03-24

By Markus Norén

DocuSigned by:

6853712A40184AF...
Quality Assurance