



CERTIFICATE OF ANALYSIS

Product: **GoBio™ Prep 16x600 40/1000 SEC** Article number: **55 430 026**

Lot number: D07012501

Storage solution: **20% ethanol**

	Measured Value	Specification
K _D (Ovalbumin)*:	0.65	0.62-0.75
Efficiency:	6141 N/m	≥5000 N/m
Asymmetry:	1.0	0.7-1.4
Bed height:	600 mm	600-620 mm

*Property measured on resin.

Resin lot number: 157541

Release date (year-month-day): 2025-01-28


Expiry date (year-month): 2029-05

Manufacturing date (year-month): 2025-01

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2025-01-28

By Markus Norén

DocuSigned by:

6853712A40184AF...
Quality Assurance