



CERTIFICATE OF ANALYSIS

Product: **GoBio™ Prep 16x600 40/1000 SEC** Article number: **55 430 026**

Lot number: **D07012502**

Storage solution: **20% ethanol**

| | Measured Value | Specification |
|------------------------------|-----------------------|----------------------|
| K _D (Ovalbumin)*: | 0.65 | 0.62-0.75 |
| Efficiency: | 6597 N/m | ≥5000 N/m |
| Asymmetry: | 1.0 | 0.7-1.4 |
| Bed height: | 600 mm | 600-620 mm |

*Property measured on resin.

Resin lot number: 157541

Release date (year-month-day): 2025-01-28

Expiry date (year-month): 2029-05

Manufacturing date (year-month): 2025-01

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2025-01-28

By Markus Norén

DocuSigned by:

Markus Norén

6853712A40184AF...

Quality Assurance