



---

## CERTIFICATE OF ANALYSIS

---

Product:  
GoBio™ Prep 16x600 40/100 SEC

Article number:  
55 434 026

Lot number: D07032501

Storage solution:

20% ethanol

---

	Measured Value	Specification
K <sub>D</sub> (Ovalbumin)*:	0.55	0.40-0.60
Efficiency:	8099 N/m	≥5000 N/m
Asymmetry:	1.4	0.7-1.4
Bed height:	600 mm	600-620 mm

\*Property measured on resin.

Resin lot number: 157833

---

Release date (year-month-day): 2025-03-24

Expiry date (year-month): 2029-11

Manufacturing date (year-month): 2025-03

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2025-03-24

By Markus Norén

DocuSigned by:

*Markus Norén*

6853712A40184AF...  
Quality Assurance