

---

## CERTIFICATE OF ANALYSIS

---

Product: **GoBio™ Prep 16x600 40/1000 SEC** Article number: 55 430 026

Lot number: D08012501

Storage solution: 20% ethanol

---

	Measured Value	Specification
K <sub>D</sub> (Ovalbumin)*:	0.65	0.62-0.75
Efficiency:	5574 N/m	≥5000 N/m
Asymmetry:	1.0	0.7-1.4
Bed height:	600 mm	600-620 mm

\*Property measured on resin.

Resin lot number: 157541

---

Release date (year-month-day): 2025-01-28

Expiry date (year-month): 2029-05


Manufacturing date (year-month): 2025-01

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2025-01-28

By Markus Norén

DocuSigned by:  
  
6853712A40184AF...  
Quality Assurance