



CERTIFICATE OF ANALYSIS

Product:
GoBio™ Prep 26x100 40Q

Article number:
55 410 031

Lot number: D09052412

Storage solution:

20% ethanol

| | Measured Value | Specification |
|------------------|------------------------------|----------------------------------|
| Ionic capacity*: | 203 µmol Cl ⁻ /ml | 180-250 µmol Cl ⁻ /ml |
| Efficiency: | 10905 N/m | ≥4000 N/m |
| Asymmetry: | 1.2 | 0.7-1.4 |

*Property measured on resin.

Resin lot number: 157485

Release date (year-month-day): 2024-05-24


Expiry date (year-month): 2029-04

Manufacturing date (year-month): 2024-05

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2024-05-24

By Markus Norén

DocuSigned by:

6853712A40184AF...
Quality Assurance