

CERTIFICATE OF ANALYSIS

Product:
GoBio™ Screen 7x100 Zn-IDA

Article number:
55 650 501

Lot number: D10072503

Storage solution:

20% ethanol

	Measured Value	Specification
Metal binding capacity before charging with Cu ²⁺ .*	52 µmol Cu ²⁺ /mL	50-70 µmol Cu ²⁺ /mL

*Property measured on resin.
Resin lot number: 158075

Release date (year-month-day): 2025-07-30


Expiry date (year-month): 2030-03

Manufacturing date (year-month): 2025-07

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2025-07-30

By Johanna Tawe

DocuSigned by:

78ADB347C12D40E...

Quality Assurance