



CERTIFICATE OF ANALYSIS

Product:
GoBio™ Screen 7x100 40Q

Article number:
55 410 001

Lot number: D10092509

Storage solution:

20% ethanol

| | Measured Value | Specification |
|------------------|------------------------------|----------------------------------|
| Ionic capacity*: | 183 µmol Cl ⁻ /ml | 180-250 µmol Cl ⁻ /ml |

*Property measured on resin.

Resin lot number: 158216

Release date (year-month-day): 2025-09-23

Expiry date (year-month): 2030-04

Manufacturing date (year-month): 2025-09

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2025-09-23

By Markus Norén

DocuSigned by:

Markus Norén

6853712A40184AF...

Quality Assurance