



CERTIFICATE OF ANALYSIS

Product:
GoBio™ Prep 16x100 40 TREN

Article number:
55 463 021

Lot number:D10122402

Storage solution:

20% ethanol

	Measured Value	Specification
Ionic capacity*:	142 µmol Cl ⁻ /ml	130-200 µmol Cl ⁻ /ml
Efficiency:	≥4000 N/m	≥4000 N/m
Asymmetry:	0.99	0.7-1.4

*Property measured on resin.

Resin lot number: 153728

Release date (year-month-day): 2025-01-13

Expiry date (year-month): 2027-04

Manufacturing date (year-month): 2024-12

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day):2025-01-13

By Markus Norén

DocuSigned by:

Markus Norén

6853712A40184AF...

Quality Assurance