
CERTIFICATE OF ANALYSIS

Product:

GoBio™ Prep 16x100 40Q

Article number:

55 410 021

Lot number: D11052201

Storage solution:

20% ethanol

	Measured Value	Specification
Ionic capacity*:	191 µmol Cl-/ml	180-250 µmol Cl-/ml
Efficiency:	4462 N/m	≥4000 N/m
Asymmetry:	1.0	0.7-1.4

*Property measured on resin.

Resin lot number: 155438

Release date (year-month-day): 2022-07-15

Expiry date (year-month): 2027-02

Manufacturing date (year-month): 2022-07

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.

Country of origin is Sweden.

Issued (year-month-day): 2022-07-15

By Anna Boman



Quality Assurance