
CERTIFICATE OF ANALYSIS

Product:
GoBio™ Prep 26x100 40Q

Article number:
55 410 031

Lot number: D11112201

Storage solution:

20% ethanol

	Measured Value	Specification
Ionic capacity*:	192 $\mu\text{mol Cl}^-/\text{ml}$	180-250 $\mu\text{mol Cl}^-/\text{ml}$
Efficiency:	6322 N/m	≥ 4000 N/m
Asymmetry:	0.94	0.7-1.4

*Property measured on resin.

Resin lot number: 155786

Release date (year-month-day): 2022-11-25


Expiry date (year-month): 2027-05

Manufacturing date (year-month): 2022-11

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2022-11-25

By Philip Eriksson

DocuSigned by:

158C686C7E5E4A0...
Quality Assurance