



CERTIFICATE OF ANALYSIS

Product: **GoBio™ Prep 16x600 40/10 000 SEC** Article number: **55 435 026**

Lot number: **D12072403**

Storage solution: **20% ethanol**

	Measured Value	Specification
K _D (Ovalbumin)*:	0.71	0.65-0.85
Efficiency:	7422 N/m	≥5000 N/m
Asymmetry:	1.2	0.7-1.4
Bed height:	600 mm	600-620 mm

*Property measured on resin.

Resin lot number: 157080

Release date (year-month-day): 2024-08-14


Expiry date (year-month): 2028-11

Manufacturing date (year-month): 2024-07

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2024-08-14

By Markus Norén

DocuSigned by:

Quality Assurance
6853712A40184AF...