



---

## CERTIFICATE OF ANALYSIS

---

Product: **GoBio™ Prep 26x600 40/10 000 SEC** Article number: **55 435 036**

Lot number: **D13012501**

Storage solution: **20% ethanol**

---

	<b>Measured Value</b>	<b>Specification</b>
K <sub>D</sub> (Ovalbumin)*:	0.71	0.65-0.85
Efficiency:	10108 N/m	≥5000 N/m
Asymmetry:	1.1	0.7-1.4
Bed height:	600 mm	600-620 mm

\*Property measured on resin.

Resin lot number: 157736

---

Release date (year-month-day): 2025-02-12

Expiry date (year-month): 2029-08


Manufacturing date (year-month): 2025-01

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2025-02-12

By Markus Norén

DocuSigned by:  
  
6853712A40184AF...  
Quality Assurance