
CERTIFICATE OF ANALYSIS

Product:
GoBio™ Prep 16x100 40S

Article number:
55 420 021

Lot number: D13102204

Storage solution:

20% ethanol, 0.2 M sodium acetate

	Measured Value	Specification
Ionic capacity*:	237 $\mu\text{mol H}^+/\text{ml}$	180–250 $\mu\text{mol H}^+/\text{ml}$
Efficiency:	4546 N/m	≥ 4000 N/m
Asymmetry:	1.2	0.7–1.4

*Property measured on resin.

Resin lot number: 156142

Release date (year-month-day): 2022-11-04

Expiry date (year-month): 2027-09

Manufacturing date (year-month): 2022-10

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2022-11-04

By Anna Boman

DocuSigned by:



3C2F4EE14561428...

Quality Assurance