



---

## CERTIFICATE OF ANALYSIS

---

Product:

GoBio™ Prep 26x100 Ni-NTA

Article number:

55 651 031

Lot number: D15072401

Storage solution:

20% ethanol

---

|  | Measured Value                 | Specification                   |
|--|--------------------------------|---------------------------------|
| Metal binding capacity before binding with Ni <sup>2+</sup> *: | 25.0 µmol Cu <sup>2+</sup> /ml | 20-30 µmol Cu <sup>2+</sup> /ml |
| Efficiency:  | 6206 N/m                       | ≥4000 N/m                       |
| Asymmetry:   | 1.2                            | 0.7-1.4                         |

\*Property measured on resin.

Resin lot number: 600064-3035-026Ni

Release date (year-month-day): 2024-08-14

Expiry date (year-month): 2026-02

Manufacturing date (year-month): 2024-07

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2024-08-14

By Markus Norén

DocuSigned by:

6853712A40184AF...

Quality Assurance