



---

## CERTIFICATE OF ANALYSIS

---

Product:

GoBio™ Prep 26x100 Ni-NTA

Article number:

55 651 031

Lot number: D15072402

Storage solution:

20% ethanol

---

	Measured Value	Specification
Metal binding capacity before binding with Ni <sup>2+</sup> *:	25.0 µmol Cu <sup>2+</sup> /ml	20-30 µmol Cu <sup>2+</sup> /ml
Efficiency:	4035 N/m	≥4000 N/m
Asymmetry:	1.2	0.7-1.4

\*Property measured on resin.

Resin lot number: 600064-3035-026Ni

Release date (year-month-day): 2024-09-17

Expiry date (year-month): 2026-02

Manufacturing date (year-month): 2024-07

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2024-09-17

By Markus Norén

DocuSigned by:

6853712A40184AF...

Quality Assurance