
CERTIFICATE OF ANALYSIS

Product:

GoBio™ Prep 16x100 40 TREN

Article number:

55 463 021

Lot number: D17052202

Storage solution:

20% ethanol

	Measured Value	Specification
Ionic capacity*:	163 µmol Cl ⁻ /ml	130-200 µmol Cl ⁻ /ml
Efficiency:	7242 N/m	≥4000 N/m
Asymmetry:	1.0	0.7-1.4

*Property measured on resin.

Resin lot number: 153892

Release date (year-month-day): 2022-07-18

Expiry date (year-month): 2025-04

Manufacturing date (year-month): 2022-07

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2022-07-18

By Anna Boman



Quality Assurance