Docusign Envelope ID: 21FD1449-2272-43A3-A703-4A71B9E323BF

Bio-Works

CERTIFICATE OF ANALYSIS

Product:

GoBio™ Prep 16x600 40/100 SEC

Article number:
55 434 026

Lot number: D18022601

Storage solution: 20% ethanol

	Measured Value	Specification
K_D (Ovalbumin)*:	0.55	0.40-0.60
Efficiency:	7068 N/m	≥5000 N/m
Asymmetry:	1.4	0.7-1.4
Bed height:	600 mm	600-620 mm

*Property measured on resin.
Resin lot number: 157883

Release date (year-month-day): 2026-03-04

Expiry date (year-month): 2029-11

Manufacturing date (year-month): 2026-02

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Manufactured in the UK, for Bio-Works, Sweden.

Issued (year-month-day): 2026-03-04
By Markus Norén

DocuSigned by:
Markus Norén
68537A2A40184AE...
Quality Assurance

Bio-Works Sweden AB
Virdings allé 18
SE 754 50 Uppsala
Sweden