



CERTIFICATE OF ANALYSIS

Product:
GoBio™ Prep 16x600 40/100 SEC

Article number:
55 434 026

Lot number: D18022602

Storage solution:

20% ethanol

	Measured Value	Specification
K _D (Ovalbumin)*:	0.55	0.40-0.60
Efficiency:	7928 N/m	≥5000 N/m
Asymmetry:	1.4	0.7-1.4
Bed height:	600 mm	600-620 mm

*Property measured on resin.
Resin lot number: 157883

Release date (year-month-day): 2026-03-04

Expiry date (year-month): 2029-11

Manufacturing date (year-month): 2026-02

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Manufactured in the UK, for Bio-Works, Sweden.

Issued (year-month-day): 2026-03-04

By Markus Norén

DocuSigned by:



Quality Assurance