

---

## CERTIFICATE OF ANALYSIS

---

Product:

GoBio™ Prep 16x100 40Q

Article number:

55 410 021

Lot number: D20072207

Storage solution:

20% ethanol

---

	Measured Value	Specification
Ionic capacity*:	192 µmol Cl <sup>-</sup> /ml	180-250 µmol Cl <sup>-</sup> /ml
Efficiency:	5460 N/m	≥4000 N/m
Asymmetry:	1.1	0.7-1.4

\*Property measured on resin.

Resin lot number: 155786

---

Release date (year-month-day): 2022-08-18

Expiry date (year-month): 2027-05

Manufacturing date (year-month): 2022-07

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.

Country of origin is Sweden.

Issued (year-month-day): 2022-08-18

By Anna Boman



Quality Assurance