



CERTIFICATE OF ANALYSIS

Product:

GoBio™ Screen 7x100 Ni-NTA

Article number:

55 651 001

Lot number: D21052501

Storage solution:

20% ethanol

	Measured Value	Specification
Metal binding capacity before charging with Ni ²⁺ .*	21 µmol Cu ²⁺ /mL	20-30 µmol Cu ²⁺ /mL

*Property measured on resin.

Resin lot number: 600064-3041-028Ni-2

Release date (year-month-day): 2025-06-10

Expiry date (year-month): 2028-06

Manufacturing date (year-month): 2025-05

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2025-06-10

By Markus Norén

DocuSigned by:

6853712A40184AF...

Quality Assurance