

---

## CERTIFICATE OF ANALYSIS

---

Product:  
GoBio™ Prep 16x100 NiMAC

Article number:  
55 653 021

Lot number: D22072204

Storage solution:

20% ethanol

---

	Measured Value	Specification
Metal binding capacity before binding with Ni <sup>2+</sup> *:	>60 µmol Cu <sup>2+</sup> /ml	>60 µmol Cu <sup>2+</sup> /ml
Efficiency:	4186 N/m	≥4000 N/m
Asymmetry:	1.3	0.7-1.4

\*Property measured on resin.

Resin lot number: 600366-3032-025N

---

Release date (year-month-day): 2022-08-22

Expiry date (year-month): 2026-04

Manufacturing date (year-month): 2022-07

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2022-08-22

By Philip Eriksson

DocuSigned by:

Philip Eriksson

158C686C7E5E4A0...  
Quality Assurance