



CERTIFICATE OF ANALYSIS

Product:
GoBio™ Prep 16x100 100Q

Article number:
55 110 021

Lot number: D23042508

Storage solution:

20% ethanol

	Measured Value	Specification
Ionic capacity*:	154 µmol Cl ⁻ /ml	140-200 µmol Cl ⁻ /ml
Efficiency:	2581 N/m	≥1500 N/m
Asymmetry:	1.2	0.7-1.4

*Property measured on resin.

Resin lot number: 157237

Release date (year-month-day): 2025-06-27

Expiry date (year-month): 2029-01

Manufacturing date (year-month): 2025-04

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2025-06-27

By Johanna Tawe

DocuSigned by:

78ADB347C12D40E...

Quality Assurance