
CERTIFICATE OF ANALYSIS

Product:
GoBio™ Prep 16x100 NTA

Article number:
55 602 021

Lot number: D23102302

Storage solution:

20% ethanol

	Measured Value	Specification
Metal binding capacity*:	25 µmol Cu ²⁺ /ml	20-30 µmol Cu ²⁺ /ml
Efficiency:	5779 N/m	≥4000 N/m
Asymmetry:	1.1	0.7-1.4

*Property measured on resin.

Resin lot number: 600064-3032-032

Release date (year-month-day): 2023-11-09

Expiry date (year-month): 2026-09

Manufacturing date (year-month): 2023-10

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2023-11-09

By Åsa Nilsson

DocuSigned by:

Åsa Nilsson

FD9074A2ACBC431...

Quality Assurance