

---

## CERTIFICATE OF ANALYSIS

---

Product:  
GoBio™ Prep 26x100 40Q

Article number:  
55 410 031

Lot number: D24022303

Storage solution:

20% ethanol

---

	Measured Value	Specification
Ionic capacity*:	201 µmol Cl <sup>-</sup> /ml	180-250 µmol Cl <sup>-</sup> /ml
Efficiency:	8953 N/m	≥4000 N/m
Asymmetry:	1.1	0.7-1.4

\*Property measured on resin.

Resin lot number: 155915

---

Release date (year-month-day): 2023-03-13

Expiry date (year-month): 2027-06

Manufacturing date (year-month): 2022-06

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2023-03-13

By Jonatan Bylund

DocuSigned by:



F0489A1BCFA8414..

Quality Assurance