
CERTIFICATE OF ANALYSIS

Product: GoBio™ Screen 7x100 Ni-NTA Article number: 55 651 001

Lot number: D24022308

Storage solution: 20% ethanol

	Measured Value	Specification
Metal binding capacity before charging with Ni ²⁺ .*	25 µmol Cu ²⁺ /mL	20-30 µmol Cu ²⁺ /mL

*Property measured on resin.


Resin lot number: 600064-3029-025

Release date (year-month-day): 2023-03-13 Expiry date (year-month): 2025-05
Manufacturing date (year-month): 2020-05

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2023-03-13

By Jonatan Bylund

DocuSigned by:

F0489A1BCFA8414...
Quality Assurance