

---

## CERTIFICATE OF ANALYSIS

---

Product:  
GoBio™ Prep 16x100 Ni-IDA

Article number:  
55 650 021

Lot number: D25082202

Storage solution:

20% ethanol

---

	Measured Value	Specification
Metal binding capacity before binding with Ni <sup>2+</sup> *:	58 µmol Cu <sup>2+</sup> /ml	50-70 µmol Cu <sup>2+</sup> /ml
Efficiency:	5249 N/m	≥4000 N/m
Asymmetry:	0.95	0.7-1.4

\*Property measured on resin.

Resin lot number: 153071

---

Release date (year-month-day): 2022-09-13

Expiry date (year-month): 2023-12

Manufacturing date (year-month): 2022-09

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2022-09-13

By Philip Eriksson

DocuSigned by:



Quality Assurance