
CERTIFICATE OF ANALYSIS

Product:
GoBio™ Prep 16x100 Ni-IDA

Article number:
55 650 021

Lot number: D25082204

Storage solution:

20% ethanol

	Measured Value	Specification
Metal binding capacity before binding with Ni ²⁺ *:	58 µmol Cu ²⁺ /ml	50-70 µmol Cu ²⁺ /ml
Efficiency:	6677 N/m	≥4000 N/m
Asymmetry:	1.1	0.7-1.4

*Property measured on resin.

Resin lot number: 153071

Release date (year-month-day): 2022-09-13

Expiry date (year-month): 2023-12


Manufacturing date (year-month): 2022-09

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2022-09-13

By Philip Eriksson

DocuSigned by:



158C888C7E5E4A0...
Quality Assurance