



CERTIFICATE OF ANALYSIS

Product:
GoBio™ Prep 16x100 NiMAC

Article number:
55 653 021

Lot number: D26082502

Storage solution:

20% ethanol

	Measured Value	Specification
Metal binding capacity before binding with Ni ²⁺ *:	110 µmol Cu ²⁺ /ml	>60 µmol Cu ²⁺ /ml
Efficiency:	4305 N/m	≥4000 N/m
Asymmetry:	1.2	0.7-1.4

*Property measured on resin.

Resin lot number: 600366-3041-026Ni

Release date (year-month-day): 2025-09-05

Expiry date (year-month): 2027-12

Manufacturing date (year-month): 2025-08

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2025-09-05

By Markus Norén

DocuSigned by:

Markus Norén

6853712A40184AF...

Quality Assurance