



---

## CERTIFICATE OF ANALYSIS

---

Product:  
GoBio™ Prep 16x100 NiMAC

Article number:  
55 653 021

Lot number: D26082503

Storage solution:

20% ethanol

---

	Measured Value	Specification
Metal binding capacity before binding with Ni <sup>2+</sup> *:	110 µmol Cu <sup>2+</sup> /ml	>60 µmol Cu <sup>2+</sup> /ml
Efficiency:	4015 N/m	≥4000 N/m
Asymmetry:	1.2	0.7-1.4

\*Property measured on resin.

Resin lot number: 600366-3041-026Ni

---

Release date (year-month-day): 2025-09-05

Expiry date (year-month): 2027-12

Manufacturing date (year-month): 2025-08

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2025-09-05

By Markus Norén

DocuSigned by:

*Markus Norén*

6853712A40184AF...

Quality Assurance