
CERTIFICATE OF ANALYSIS

Product:
GoBio™ Prep 26x100 Ni-IDA

Article number:
55 650 031

Lot number: D27092201

Storage solution:

20% ethanol

| | Measured Value | Specification |
|----------------------------------------------------------------|------------------------------|---------------------------------|
| Metal binding capacity before binding with Ni ²⁺ *: | 64 µmol Cu ²⁺ /ml | 50-70 µmol Cu ²⁺ /ml |
| Efficiency: | 6019 N/m | ≥4000 N/m |
| Asymmetry: | 1.4 | 0.7-1.4 |

*Property measured on resin.

Resin lot number: 154855

Release date (year-month-day): 2022-10-07


Expiry date (year-month): 2026-07

Manufacturing date (year-month): 2022-10

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2022-10-07

By Philip Eriksson

DocuSigned by:

158C886C7E5E4A0...
Quality Assurance