
CERTIFICATE OF ANALYSIS

Product:
GoBio™ Prep 26x100 Ni-IDA

Article number:
55 650 031

Lot number: D27092202

Storage solution:

20% ethanol

	Measured Value	Specification
Metal binding capacity before binding with Ni ²⁺ *:	64 µmol Cu ²⁺ /ml	50-70 µmol Cu ²⁺ /ml
Efficiency:	4595 N/m	≥4000 N/m
Asymmetry:	1.4	0.7-1.4

*Property measured on resin.

Resin lot number: 154855

Release date (year-month-day): 2022-10-06

Expiry date (year-month): 2026-07


Manufacturing date (year-month): 2022-10

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2022-10-06

By Philip Eriksson

DocuSigned by:



158C886C7E5E4A0...

Quality Assurance