
CERTIFICATE OF ANALYSIS

Product: GoBio™ Prep 26x100 Ni-NTA Article number: 55 651 031

Lot number: D28042302

Storage solution: 20% ethanol

	Measured Value	Specification
Metal binding capacity before binding with Ni ²⁺ *:	25 µmol Cu ²⁺ /ml	20-30 µmol Cu ²⁺ /ml
Efficiency:	6221 N/m	≥4000 N/m
Asymmetry:	1.2	0.7-1.4

*Property measured on resin.

Resin lot number: 600064-3029-025

Release date (year-month-day): 2023-05-15


Expiry date (year-month): 2025-05

Manufacturing date (year-month): 2020-05

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2023-05-15

By Philip Eriksson

DocuSigned by:

158C686C7E5E4A0...
Quality Assurance