
CERTIFICATE OF ANALYSIS

Product:
GoBio™ Prep 16x600 Macro SEC

Article number:
55 437 026

Lot number: D29022402

Storage solution:

20% ethanol

	Measured Value	Specification
K _D (Ovalbumin)*:	0.78	0.70-0.90
Efficiency:	8420 N/m	≥5000 N/m
Asymmetry:	0.83	0.7-1.4
Bed height:	600 mm	600-620 mm

*Property measured on resin.

Resin lot number: 156432

Release date (year-month-day): 2024-03-14


Expiry date (year-month): 2028-01

Manufacturing date (year-month): 2024-02

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2024-03-14

By Åsa Nilsson

DocuSigned by:

FD9074A2ACBC431...
Quality Assurance