



---

## CERTIFICATE OF ANALYSIS

---

Product:  
GoBio™ Prep 16x100 40 DEAE

Article number:  
55 415 021

Lot number: D30082403

Storage solution:

20% ethanol

---

	Measured Value	Specification
Ionic capacity*:	136 $\mu\text{mol Cl}^-/\text{ml}$	110-160 $\mu\text{mol Cl}^-/\text{ml}$
Efficiency:	6112 N/m	$\geq 4000$ N/m
Asymmetry:	1.2	0.7-1.4

\*Property measured on resin.

Resin lot number: 156725

---

Release date (year-month-day): 2024-09-24

Expiry date (year-month): 2028-05

Manufacturing date (year-month): 2024-08

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2024-09-24

By Markus Norén

DocuSigned by:

*Markus Norén*

6853712A40184AF...

Quality Assurance