
CERTIFICATE OF ANALYSIS

Product: GoBio™ Prep 16x600 40/1000 SEC Article number: 55 430 026

Lot number: D31012310

Storage solution: 20% ethanol

	Measured Value	Specification
K _D (Ovalbumin)*:	0.64	0.62-0.75
Efficiency:	7324 N/m	≥5000 N/m
Asymmetry:	0.91	0.7-1.4
Bed height:	600 mm	600-620 mm

*Property measured on resin.

Resin lot number: 155743

Release date (year-month-day): 2023-02-21

Expiry date (year-month): 2027-05


Manufacturing date (year-month): 2023-02

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2023-02-21

By Philip Eriksson

DocuSigned by:



158C686C7E5E4A0...

Quality Assurance