

## CERTIFICATE OF ANALYSIS

---

Product: **WorkBeads™ 40/100 SEC**

Article number: **40 340 001, 40 340 003, 40 340 010, 40 340 050,  
40 340 250, 40 340 260**

Lot number: **153936**

Product release date: **2020 JUN 02**

Expire date: **2023-06**


---

	Measured values	Specifications
Average particle size (Dv50):	54 µm	32 - 60 µm
Kd (Ovalbumin):	0.55	0.40 - 0.60
Flow properties (linear flow in a 10x300 mm column):	>10 cm/min	≥10 cm/min

---

Issued **2020 JUN 02**

by



---

Johanna Tawe  
QMS Director