

CERTIFICATE OF ANALYSIS

Product: BabyBio Q 5 ml

Article number: 45 100 105, 45 100 106, 45 100 107, 45 100 108

Lot number: 152490

Resin lot number: 151946

	Measured values	Specifications
Ionic capacity:	180 µmol Cl/ml	180-250 µmol Cl/ml
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued 2019 Apr 01

by



Mats Johnson
Chief Executive officer