

CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 40/1000 SEC**

Article number: **40 300 001, 40 300 003, 40 300 010, 40 300 050,
40 300 60, 40 300 201, 40 300 203, 40 300 250,
40 300 260**

Lot number: **152592**

Manufacturing date: **2018 APRIL 20**

	Measured values	Specifications
Average particle size (Dv50)	51 µm	32 - 60 µm
Kd (Ovalbumin)	0.65	0.62 - 0.75
Flow properties (linear flow in a 10x300 mm column)	> 20 cm/min	≥10 cm/min

Issued 2018 JUNE 15

by



Mats Johnson
Chief Executive officer