
CERTIFICATE OF ANALYSIS

Product: BabyBio NTA His-tag Screening kit 1 ml

Article number: 45 700 101

Lot number: 153310

Resin lot numbers:


BabyBio Ni-NTA 1ml: 152466
BabyBio Cu-NTA 1ml: 151530
BabyBio Co-NTA 1ml: 152062
BabyBio Zn-NTA 1ml: 151546

Product release date: 2016 JAN 08
Expire date: 2021-01

	Measured values	Specifications
Metal ion capacity (Tested with Cu ²⁺)		20 – 30 µmol/ml
BabyBio Ni-NTA 1 ml	23 µmol/ml	
BabyBio Cu-NTA 1 ml	24 µmol/ml	
BabyBio Co-NTA 1 ml	24 µmol/ml	
BabyBio Zn-NTA 1 ml	24 µmol/ml	
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued **2020 MAY 29**

by



Helena Pahlm
Quality Assurance Manager