



STATEMENT OF ANALYSIS

Product:	Article number:
GoBio™ Validation 16x140 DMAE	55 180 001
Lot number: D20072301	
Storage solution:	20% ethanol

	Measured Value
Ligand concentration*:	104 µmol/mL
Efficiency:	6182 N/m
Asymmetry:	1.1
Delta column pressure at 8 mL/min	0.11 Mpa
Bed height	14.0 cm

*Property measured on resin.
Resin lot number: 073060_C_L

Issued (year-month-day): 2023-08-04

By Xiao Huang

DocuSigned by:

A49E9003037D4EF...
Product Manager