

CERTIFICATE OF ANALYSIS

Product:	BabyBio DEAE 5 ml
Storage solution:	20% ethanol
Article number:	45 150 105, 45 150 106, 45 150 107, 45 150 108
Lot number:	154343
Resin lot number:	153821
Manufacturing date:	2021 JAN 12
Product release date:	2021 JAN 21
Expire date:	2026-01

	Measured values	Specifications
Ionic capacity:	142 $\mu\text{mol Cl}^-/\text{ml}$	110-160 $\mu\text{mol Cl}^-/\text{ml}$
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued
by



Afifa Trad
Quality Assurance Director