

CERTIFICATE OF ANALYSIS

Product:	BabyBio TREN 1 ml
Storage solution:	20% ethanol
Article number:	45 655 211, 45 655 212, 45 655 213, 45 655 214
Lot number:	154344
Resin lot number:	153775
Manufacturing date:	2021 JAN 15
Product release date:	2021 JAN 21
Expire date:	2026-01

	Measured values	Specifications
Metal ion capacity (Tested with Cu ²⁺)	51 µmol Cu ²⁺ /ml	50 – 60 µmol Cu ²⁺ /ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued
by



Afifa Trad
Quality Assurance Director

