

CERTIFICATE OF ANALYSIS

Product: BabyBio Q 1 mL

Article number: 45 100 101, 45 100 102, 45 100 103, 45 100 104

Lot number: 151265

Resin lot number: BB 007

	Measured values	Specifications
Ionic capacity	190 $\mu\text{mol Cl}^-/\text{mL}$	180-250 $\mu\text{mol Cl}^-/\text{mL}$
Gap control: (5 mL/min, water, 22 °C)	Passed	No gap
Maximum flow test: (7 mL/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued 2018 AUGUST 16

by



Mats Johnson
Chief Executive officer