

## CERTIFICATE OF ANALYSIS

---

Product: **WorkBeads™ 40 Co-IDA**

Article number: **40 650 401, 40 650 403, 40 650 410**

Lot number: **151401**

Manufacturing date: **2016 FEB 29**


---

	<b>Measured values</b>	<b>Specifications</b>
Capacity before charging with Co <sup>2+</sup>	67 µmol Cu <sup>2+</sup> /ml	50 - 70 µmol Cu <sup>2+</sup> /ml
Average particle size (Dv50):	49.6 µm	32 - 60 µm
Flow acceptance (linear flow in a 10x300 mm column):	>10 cm/min	≥10 cm/min

---

Issued 2019 FEB 22

by



**Mats Johnson**  
Chief Executive officer