

---

## CERTIFICATE OF ANALYSIS

---

<b>Product:</b>	<b>BabyBio Zn-IDA 5 ml</b>
Article number:	<b>45 655 045, 45 655 046, 45 655 047, 45 655 048</b>
Lot number:	<b>151525</b>
Resin lot number:	<b>151417</b>
Product release date:	<b>2015 DEC 04</b>
Expire date:	<b>2020-12</b>

---

	<b>Measured values</b>	<b>Specifications</b>
Capacity before charging with Zn <sup>2+</sup>	67 µmol Cu <sup>2+</sup> /ml	50 – 70 µmol Cu <sup>2+</sup> /ml
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

---

Issued **2020 MAY 29**

by



Helena Pahlm  
Quality Assurance Manager