
CERTIFICATE OF ANALYSIS

Product: BabyBio Zn-NTA 1 ml

Article number: 45 655 141, 45 655 142, 45 655 143, 45 655 144

Lot number: 151546

Resin lot number: 151426

Product release date: 2016 JAN 08

Expire date: 2021-01

	Measured values	Specifications
Colour:	Passed	White to off-white
Metal ion capacity (Tested with Cu ²⁺)	24 µmol/ml	20 – 30 µmol/ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued **2020 MAY 07**

by



Mats Johnson
Chief Executive officer